



Republic of the Philippines
Supreme Court
Manila

**BIDS AND AWARDS COMMITTEE FOR GOODS AND SERVICES
(SC-BAC-GS)**

BID BULLETIN No. 1

*Procurement of Clinical Chemistry Analyzer
for the Official Use of the Supreme Court*

Notice is hereby given of the revised technical specifications, upon consultation with the end-user / implementing office representatives:

Section VII. Technical Specifications

ITEM & DESCRIPTION		Minimum Requirements
1	Test Panels	Basic Metabolic Panel, Comprehensive Metabolic Panel, Hepatic Panel, Lipid Panel, Renal Panel
2	Test Throughput (Theoretical Max)	65 tests/hr for Photometry Tests
3	Reagent Slots:	27 slots per reagent disk, up to 8 disks
4	Programmable Parameters	Maximum 25 photometric with software ver 1.0, 10 profiles
5	Sample Types	Serum, Plasma, Urine, Whole Blood
6	Sample Input	Load capacity: 8 positions, continuous loading, stat: real stat interrupt is available, continuous operation: additional sample setting during operation
7	Photometer	12 wavelengths, 20 W halogen lamp, bichromatic measurement
8	Water/Waste Requirements	Up to 3 litres/day NCCLS Type II, biohazards waste: separate container
9	Warranty on Parts	One (1) Year
10	Service Warranty	One (1) Year Free Service Warranty
11	Training	The bidder must provide training for the operation of the Clinical Chemistry Analyzer



12	Scope of Work and Schedule of Requirements	Complete supply, delivery, installation and calibration of one (1) unit of Clinical Chemistry Analyzer, within a period of thirty (30) calendar days from receipt of the Purchase Order (manual or electronic).
----	--	---

Prospective bidders are advised to get a copy of the bid bulletin and the revised bidding form/s from the Secretariat.

Please be guided accordingly.

17 April 2023

LILIAN C. BARRIBAL-CO
Assistant Court Administrator
and Chairperson, SC-BAC-GS